



Metropolis Healthcare Ltd., EQAS
SRI SRI COLLEGE OF AYURVEDIC SCIENCE AND RESEARCH HOSPITAL
(1351)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 02

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
* Total Leucocytes (WBC) count	Sysmex XP-100	8.50	/c.mm	*
# Lymphocytes Differential	Sysmex XP-100	1.10	%	0
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⊖ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Sysmex XP-100

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.66	2.66	0.44	0.79
* Total Leucocytes (WBC) count	/c.mm	8.50	8847.83	*	0.00
@ Neutrophils Differential	%	77.30	84.18	-0.95	-2.28
# Lymphocytes Differential	%	1.10			-0.86
@ Haemoglobin	g/dL	8.00	8.11	0.55	0.42
@ MCV (Mean Corpuscular Volume)	fL	121.40	121.27	-0.60	-0.63
@ MCH (Mean Corpuscular Hb)	pg	30.10	30.72	-0.84	-1.09
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	24.80	25.44	0.46	0.13
@ Platelet Count	1000/ μ L	244.00	236.86	0.03	-0.07
@ Hematocrit	%	32.30	32.01	0.12	0.40
@ RDW CV%	%	22.40	23.00	-0.24	-0.43

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score \geq 3 - Unacceptable

: Not Evaluated

⊖ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	11
Not Evaluated Parameters	1
Evaluated Parameters	10
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	90.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070