



Metropolis Healthcare Ltd., EQAS
SRI SRI COLLEGE OF AYURVEDIC SCIENCE AND RESEARCH HOSPITAL
(1351)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 01

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit
X C-Reactive Protein Qualitative	Standard F 200	Normal	mg/L
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ ASO Qualitative	IU/mL	Normal	Normal
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Instrument : Standard F 200

Analyte	Standard Unit	Result Value	Accepted Value
X C-Reactive Protein Qualitative	mg/L	Normal	Abnormal

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	93.33 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070