



Alere Medical Pvt. Ltd.  
No. 404, BPTP Park Centre,  
NH - 8, Sector - 30  
Opp. 32nd Milestone, Gurugram  
Haryana 122001, India  
T: +91 124 4569000  
F: +91 124 4569011

## CALIBRATION CERTIFICATE

NAME OF THE INSTRUMENT : URINE CHEMISTRY ANALYSER  
MODEL NO. : UROMETER120  
SERIAL NO. : 12211882  
USER NAME : LUPIN DIAGNOSTIC LTD.  
ADDRESS : JP MEMORIAL HOSPITAL 58 MAHATMA GANDHI MARG, PRAYAGRAJ-211002

DATE OF CALIBRATION : 30/07/ 2024

CALIBRATION MATERIAL	CALIBRATION RANGE	RESULT
1-URO COLOR WHITE STRIP	333+/- 10	OK

Q.C. MATERIAL FOR CALIBRATION CHECK : RESULT

1- URO CONTROL LOW : OK

2- URO CONTROL HIGH : OK

For **Alere Medical Pvt. Ltd.**

( ABHISHEK KUMAR TIWARY)

**ENGINEER – Technical Services**