



**Installation Certificate for Standard F 200**

This is to certify that Standard F 200 Serial No FA20FAITG0107 is successfully installed and commissioned at Lupin Diagnostic, Prayagraj. The installation protocol/checklist has been successfully completed for the above instrument.

For **SD Biosensor Healthcare Pvt. Ltd.**



Aamir Ayaz Khan

Assistant Product Manager

**Installation Report****Customer Details****Instrument Details**

Name: Lupin Diagnostic

Model: Standard F 200

Contact Person(s) : Mr Pankaj Kumar

Serial Number : FA20FAITG0107

Contact No : +91 9910622970

Installation Date : 24-07-2023

Address : Inside Dr Rekha Shrivastava Hospital, Prayagraj.

**Accessories and Consumable Supplied**

Power adapter: Yes

Calibration set: Yes

Test kit, Buffer: Yes

User instruction guide: Yes

**Installation Checks**

Environment: Closed

Power Adapter: Working

Temperature: 26 degree Celsius (in limits between 15 degree to 32 degree Celsius)

UPS: Yes

**Training Imparted****Assays Performed**

Operating Procedure


Maintenance

Dos and Don'ts

Calibration

**Customer Remarks****Installer's Remarks**

Customer's Signature




Installer's Signature

### **Operations Qualification for Standard F 200**

1. Verified Cal 1 Calibration (blank check) : Done
2. Verified Cal 2 Calibration (UV light source and detector check) : Done
3. Verified Cal 3 Calibration (RBG light source and three level band Intensity (Strong positive, weak positive and negative) check) : Done

For **SD Biosensor Health Care Pvt. Ltd.**



Aamir Ayaz Khan

Assistant Product Manager

**Performance Certificate – Standard F 200**



**Instrument Certification:**

Study data has determined that the instrument described in this document either meets all criteria outlined in this Performance Qualification Protocol or exceptional conditions have been identified and documentation included.

It is recommended to do Instrument calibration by calibration set once in 30 days for checking optical parts of analyzer. There is no preventive maintenance required in this equipment. SD Biosensor will update software of system free of cost as when required to run machine at optimum levels.

This instrument is ready for specific usage.

Protocol Performed By:



Aamir Ayaz Khan

Assistant Product Manager