



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, RAJPURA (1053)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____


Reviewed by: _____

Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	5,5,5,6
@ Specific Gravity	-	1.020	1.020,1.025,1.030
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,+3
@ Ketone (urine)	-	Negative	+1,Neg,Tr
@ Blood (urine)	-	2+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	+1,Neg
@ Urobilinogen (urine)	-	Normal	+1,Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
@ : Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, RAJPURA (1053)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____


Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value										
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal										
Legend @ : Acceptable @ : Acceptable X : Unacceptable # : Not Evaluated Ⓞ : Delayed Result Entry		<table border="1"> <tr> <td>Total Parameters</td> <td>1</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>0</td> </tr> <tr> <td>Evaluated Parameters</td> <td>1</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score Clinical Path - Urine Sediment Analysis Image</td> <td>100.00 %</td> </tr> </table>		Total Parameters	1	Not Evaluated Parameters	0	Evaluated Parameters	1	Outlier Parameters (X)	0	EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %
Total Parameters	1												
Not Evaluated Parameters	0												
Evaluated Parameters	1												
Outlier Parameters (X)	0												
EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %												



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroil Road, Kurla (W),
 Mumbai - 400070