



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, RAJPURA (1053)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240207 and Sample No 02

Report Date : 18/10/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	77.70	79.72	0.05	0.05
@ Urea	mg/dl	27.80	29.05	-0.15	0.21
@ Creatinine	mg/dl	1.10	1.20	-1.10	-0.35
@ Bilirubin Total	mg/dl	0.20	0.43	-1.21	-0.12
@ Bilirubin Direct	mg/dl	0.10	0.14	-0.06	0.37
@ Total Protein	g/dL	7.40	7.41	0.09	0.39
@ Albumin	g/dL	3.90	4.09	-0.36	-2.71
@ Uric Acid	mg/dl	4.39	4.66	-0.60	-1.42
@ Calcium	mg/dl	8.41	8.77	-0.74	4.60
@ Cholesterol	mg/dl	149.00	162.70	-0.86	-0.37
@ Triglycerides	mg/dl	120.00	130.79	-1.06	-0.54
@ HDL	mg/dl	34.00	32.82	-0.20	0.21
@ LDL	mg/dl	84.82	86.09	-0.45	-0.09
@ ALT	U/L	5.30	5.21	0.61	0.84
@ AST	U/L	8.93	15.82	-1.04	0.54
@ ALP	U/L	33.03	45.79	-0.73	-0.23

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	101.24	101.75	-0.34	0.53
@ Sodium	mmol/L	127.07	129.19	-0.89	-0.93
@ Potassium	mmol/L	4.26	4.34	-0.72	-0.56

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	19
Not Evaluated Parameters	0
Evaluated Parameters	19
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070