



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, RAJPURA (1053)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! MCH (Mean Corpuscular Hb)	Horiba ABX Micros ES 60	29.70	pg	2.75
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____ **Dated:** _____

Instrument : Horiba ABX Micros ES 60

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.50	2.46	0.03	0.06
@ Total Leucocytes (WBC) count	/c.mm	4300.00	4473.85	-0.36	0.77
@ Haemoglobin	g/dL	7.40	7.09	0.76	0.75
@ MCV (Mean Corpuscular Volume)	fL	99.00	104.16	-1.12	-0.57
! MCH (Mean Corpuscular Hb)	pg	29.70	28.63	2.75	1.02
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	30.10	26.98	1.96	1.28
@ Platelet Count	1000/ µL	138.00	150.31	-0.19	-0.56
@ Hematocrit	%	24.60	26.60	-0.94	-0.43
@ RDW CV%	%	17.40	19.22	-0.33	0.30

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070