

THECNICAL CALIBRATION REPORT

To,
Hoffen Diagnostics, Pune.

Date: 01 -05 -2024

Dear Madam/Sir,

As desired by you please find herewith the Calibration report

1. Year of manufacturing : H360 , Sr.no. K10012108050
2. Date of Calibration : 01-05- 2024
3. Next due date of calibration: 30-04-2025

Test Voltages / Test Sensors -

Test Voltage	Actual Voltage	Ok / Not Ok
Logic 5 V	5.0V	OK
Motor Voltage (23 – 25V)	23.5V	OK
Pressure (2.2 kg/cm ²)	2.21Kg/cm ²	OK
Vacuum 250Hg	250Hg	OK
Clog Voltage(100V)	101V	OK

Thanking you and assuring of our prompt attention at all times.


Yours faithfully,
For TRANSASIA BIO-MEDICALS LTD.

Amrut Kulkarni
Sr. Area Service Manager

