



GLOBAL TECHNICAL SERVICES

Sec. No.25, Plot No.49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune - 411044

Email : globaltechnical007@gmail.com

Mob : 9921239827 / 7276302207 / 9028888728



CC-4048

CALIBRATION CERTIFICATE

1.CUSTOMER :-	Page No. :- 1 of 1
HOFFEN DIAGNOSTICS	SRF No :- GTS/241017/05
BAVDHAN, PUNE	Certificate No. :- GTS/241017/05-003
Ambient Temp. (°C) :- 25± 4	Date of Received :- 17.10.2024
Relative Humidity (%RH) :- 30 to 75	Date of Calibration :- 17.10.2024
Location of calibration :- In Lab	Next Calibration Due On :- 16.10.2025
Condition of Item :- Ok	Issue Date :- 21.10.2024
	Calibration method No. :- TH-WI-05
	ULR No :- CC404824000003280F

2. Description of Item	
Name :- Digital Hygrometer	Range :- -50 to 80°C/10 to 99%RH
Id No :- HOF - 13	Least Count :- 0.1°C/1%RH
Make :- APTECH HTC	Location :- LAB
Type :- Digital	Dept. :- Pathology
	Sr No :- HTC 1

3.Details of Equipment used for calibration				
Name	Certificate No.	Certified By	ID No.	Calibration Validity
Temperature and Humidity Indicator With Sensor	CAL/23-24/CC/642-5	Nashik Engineering Cluster	GTS/THIS/01	03.11.2024

***Discipline Thermal Calibration (Specific heat & Humidity)**


4.Calibration Results :-

Temperature @ 50%RH				
Calibration Points °C	Standard Reading °C	UUC Reading °C	Error in °C	Expanded Unc. in ± °C
10	9.77	10.1	0.33	0.90
20	19.81	20.2	0.39	0.90
40	39.86	40.2	0.34	0.90

Humidity @25°C				
%RH	%RH	%RH	%RH	in ± %RH
20	19.73	20	0.27	2.40
50	49.79	51	1.21	2.40
90	89.85	91	1.15	2.40

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95.45% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Global Technical Services, Pune".
- 6)The Instruments used for calibration are traceable to National/International standards and their Calibrations are valid.

Calibrated By

 Calibration Engineer
 Rohit L






Approved By

 Technical Manager
 Swapnil Bhaqwat



"Sahastrarashmi", C-10, MIDC, Ambad, Nashik-422 010. Tel.: +91 253 6699231,32,91 TeleFax: + 91 253 6699222
E-mail : calibration@nec.org.in, quality.lab@nec.org.in, testinglab@nec.org.in, info@nec.org.in, website: www.nec.org.in

CALIBRATION CERTIFICATE				Certificate No. CAL/23-24/CC/642-5	
Dig.Temp. & Humidity Indicator with Sensor				ULR No. CC224823000004938F	
				Date of Issue 05.11.2023	
Date of Calibration 04.11.2023		Next Calibration Due Date 03.11.2024		Page No. 1	No. of Pages 1
Calibrated For M/S. GLOBAL TECHNICAL SERVICES. 25, Plot No. 49/3, L.I.G. Colony, Pradhikaran, Nigdi, Pune-411044		Date of Receipt 04.11.2023		Challan No. 210	In Lab 04.11.2023
Condition of the Instrument Functional		SRF No and date CAL/23-24/CSRF/642		Calibrated at SRF Date	Department
NEC ID No. CAL/23-24/ID/642-5					
Details of Test Instrument					
Description	Dig.Temp. & Humidity Indicator with Sensor	Make	Polltech Instruments Pvt. Ltd.	Model	MIH-PM-RH-T1B-TT
Type	---	ID. No.	GTS/THIS/01	Sr No.	---
Range	-40 To 60 °C/ 0 To 100 % RH	Least Count	0.01°C / 0.01% RH	Ratio	
Accuracy	±1°C & ± 5% R.H	Calibration Procedure No.		WI/NEC/CAL/THERM/05	
Calibration Environments		Temperature	25 ± 4°C	Relative Humidity	30 to 75% RH
Reference Standard Used For Calibration (Traceable To National / International Standards)					
Description	Make/Model	Sr No.	Cal.Cert No	Valid upto	Traceability
Humidity Generator	Thunder Scientific/1200	0905052	TSC/22-23/13965-2	15.12.2023	Transcal,Banglore
Calibration Results Discipline:-Thermal					
Parameter: Temperature @ 50% RH					
Set Points	Standard Reading	Reading On UUC*	Deviation	(±) Expanded Uncertainty	
°C	°C	°C	°C	°C	
5	5.01	4.80	-0.21	0.52	
30	29.99	29.60	-0.39	0.52	
50	49.98	49.20	-0.78	0.52	
Parameter: Humidity @ 25 °C					
Set Points	Standard Reading	Reading On UUC*	Deviation	(±) Expanded Uncertainty	
%	%	%	%	%	
10	10.02	10.00	-0.02	1.21	
20	20.01	20.00	-0.01	1.21	
60	59.95	60.00	0.05	1.21	
95	95.82	95.00	-0.82	1.21	
UUC*:- Unit Under Calibration					
Remarks:					
1) Due date of calibration (1 Year) is mentioned in the certificate is as per customer request.					
2) The calibration certificate pertains to the above equipment calibration.					
3) The calibration certificate shall not be reproduced except in full,without written approval of the laboratory.					
4) Uncertainty has been calculated for a coverage factor k=2 corresponding to approximately 95.45 % Confidence Level.					
5) The Standard maintained are traceable to National / International Standard through accredited Laboratories.					
6) The observations reported represent values at the time of the measurements, and under the stated conditions. they do not convey any long term stability information.					
7) Measured O/P is average of 5 reading.					
 Prepared By Kanchan Phadol (Calibration Engineer)				 Reviewed & Approved By Rahul Golesar (Quality Manager)	
*** End of Certificate ***					

QR/NEC/CC/01