



Calibration Certificate

Customer Name	Krsnaa Diagnostics Pvt . Ltd.
Address	Inside A.P Jain Civil Hospital, Rajpura (Punjab)-140401
Contact Person	Mr. Mahinder Singh
Contact No.	+916283-858912
Instrument	Electrolyte Analyzer
Model	Acculyte-3P
Serial No.	DJ05D21080024
Instrument Status	Under Warranty
Calibration Done On	12-05-2024
Next Calibration Due On	11-05-2025

For Aspen Diagnostics Pvt. Ltd.

Nitish Mahajan
Service engineer




Rustam Ansari
Area Service Manager

Calibration Rawa Data

CALIBRATE A

27

K	=	4.00	mmol/L	64.04	mV
Na	=	140.00	mmol/L	74.12	mV
Cl	=	100.00	mmol/L	91.58	mV

----- DONE ! -----

CALIBRATE B

27

K	=	8.00	mmol/L	80.68	mV
Na	=	110.00	mmol/L	67.85	mV
Cl	=	70.00	mmol/L	97.79	mV

----- Washing . . . -----