



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/THQH/CEN/D01	Date of Issue	: 23/11/2020
		Calibration Done On	: 21/11/2020
		Calibration Due On	: 20/11/2021

Customer Name	TALUK HEAD QUARTERS HOSPITAL, HARIPAD.
---------------	--

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	-	-	68041234	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
--------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Set Reading	Measured Reading
1	1048 rpm
2	1415 rpm
3	1818 rpm
4	2161 rpm
5	3445 rpm

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By

Prabin KP (Service Engineer)

Approved By

Binoj Thomas (Quality Manager)